





# **PREFACE**

#### **MEDICAL SAFETY** POWER SOLUTIONS ENGINEERED BY MINMAX

MINMAX has an extensive history and experience with the design of medical safety DC-DC converters and AC-DC power supplies for demanding applications in both medical and healthcare instrumentation that requires medical safety and a reinforced insulation system. MINMAX offers a large standard range of high-isolation and reinforced insulated medical safety DC-DC converters with a power ranging from 1 to 20 W and AC-DC power supplies with a power ranging from 24 to 60 W.

Given the requirements for medical/healthcare applications for I/O isolation, MINMAX Medical Safety Power Solutions are rated from 3000 to 5000 VAC and possess reinforced insulation and a low leakage current for operator protection (2xMOOP) or patient protection (2xMOPP).

All medical safety products meet the latest medical safety standards (ANSI/AAMI ES 60601-1 and IEC/EN 60601-1 3<sup>rd</sup> edition) and are approved for nominal working voltages of 250 Vrms or higher.

MINMAX medical safety DC-DC converters and AC-DC power supplies offer cost-effective power solutions for demanding medical and healthcare applications in dental chairs, oral care equipment, infusion pumps, medical assist devices, medical oxygen monitors, medical carts, CT scanning, ultrasound, and many pieces of medical auxiliary equipment.

## CONTENTS

#### MEDICAL SAFETY INTRODUCTION AND APPLICATIONS

History of the Editions of Medical Safety Standard IEC 60601	US
Changes from the 2 <sup>nd</sup> Edition to the 3 <sup>rd</sup> Edition of the IEC 60601-1 Standard	05-06
solation Voltage and Safety Distance in the 3 <sup>rd</sup> Edition of the IEC 60601-1 Standard	06-07
Leakage Current in the 3 <sup>rd</sup> Edition of the IEC 60601-1 Standard	07
Insulation Type	08
Structure of the Medical Insulation System	08-09
Medical EMC Standards ————————————————————————————————————	09-10
Cost vs. Risks to Safety	10
HIGHLIGHTED PERFORMANCE OF MEDICAL SAFETY PRODUCTS ————————————————————————————————————	11-16
MINMAX MEDICAL SAFETY POWER SOLUTIONS	
Successful Application	17
Medical Safety Products Overview	18
Medical Safety 1-20W DC-DC Converters	19-24
Medical Safety 24-60W AC-DC Power Supplies	25

## MDEICAL SAFETY INTRODUCTION **AND APPLICATIONS**

#### HISTORY OF THE EDITIONS **OF MEDICAL SAFETY STANDARD IEC 60601**

- Medical Safety Standard IEC 60601 was first published in 1977; the internationally accepted and approved IEC 60601 standard is the fundamental document that addresses many risks and has been developed to help alleviate safety issues associated with electrical medical and healthcare equipment.
- The standard consists of four distinct parts: the base standard (60601-1), collateral standards (60601-1-x), particular standards (60601-2-x), and performance standards (60601-3-x).
- The base standard, IEC 60601-1, has been adopted in most major countries as a national standard.
- The 3<sup>rd</sup> edition of the approved IEC 60601-1 medical safety standard was first published by the IEC in 2005 (IEC 60601:2005). The IEC 60601 standard was adopted by the European Union in 2006 and published as EN 60601-1:2006. The 3<sup>rd</sup> edition standard was also published in 2006 by the USA but is different from the 2<sup>nd</sup> edition published by UL. The 3<sup>rd</sup> edition was published by the American Association for Medical Instrumentation (AAMI) and published as ANSI/AAMI ES 60601:2006. Canada published this medical safety standard in 2008 as CAN/ CSA 60601:2008.

The 3<sup>rd</sup> edition of Medical Safety Standard IEC 60601 went into effect at different times depending on region.

1st EDITION published by IEC

2<sup>nd</sup> EDITION published by IEC

3<sup>nd</sup> EDITION published by IEC



#### HISTORY OF THE EDITIONS OF **MEDICAL SAFETY STANDARD IEC 60601**


- In Europe, the 2<sup>nd</sup> edition of the IEC 60601-1 standard was withdrawn; all products including both new products introduced to the market and products already on sale need to be certified under the 3<sup>rd</sup> edition of the EN 60601-1 standard before the cessation date. The original cessation date for the 2<sup>nd</sup> edition in the United States is July 1, 2013. The FDA announced an extension to give US medical device designers a slight reprieve, setting the updated transition date to December 31, 2013, and the effective date for the released 3rd edition was January 1, 2014. Unlike the EU, the FDA only requires that new products after this date need to be certified to the ANSI/AAMI ES 60601-1 standard; existing products do not.
- In Canada, the original cessation date for the 2<sup>nd</sup> edition was delayed, as in the USA, with an updated transition date of June 30, 2014. The effective date for the released 3<sup>rd</sup> edition was July 1, 2014. However, only new products after this date need to satisfy the 3<sup>rd</sup> edition.
- All MINMAX medical safety AC-DC power supplies and medical safety DC-DC converters have been certified to the approved 3<sup>rd</sup> edition standard (with twice the means of protection for the majority of power supplies) and also tested against the 2<sup>nd</sup> edition.also tested against the 2<sup>nd</sup> edition.

#### CHANGES FROM THE 2<sup>nd</sup> EDITION TO THE 3<sup>rd</sup> **EDITION OF THE IEC 60601-1 STANDARD**

- In the 2<sup>nd</sup> edition of the IEC 60601-1 standard, the guidelines need to be applied when the electrical medical and healthcare equipment was within the "patient vicinity," defined as a 6-ft. radius around the patient.
- There were three categories of increasing severity:also tested against the 2 nd edition.



#### Type.CF | cardiacfloating



#### CHANGES FROM THE 2<sup>nd</sup> EDITION TO THE 3<sup>rd</sup> **EDITION OF THE IEC 60601-1 STANDARD**

- The 2<sup>nd</sup> edition of the IEC 60601-1 standard only addresses the basic safety issues for protection from any electrical, mechanical, radiation, and thermal hazards. However, it did not require devices to remain functional; a fail-safe was adequate and did not take the essential performance of the device into account in a pass/fail result test. Therefore, the 3<sup>rd</sup> edition of the IEC 60601-1 standard introduces specifications for "essential performance" that require medical/healthcare equipment to continue functioning throughout the test process.
- As medical safety AC-DC power supplies and medical safety DC-DC converters have significantly played a crucial role in certified electrical medical and healthcare equipment, the 3<sup>rd</sup> edition of the IEC 60601-1 standard introduces new concepts such as the essential performance of equipment and distinguishes protecting the equipment's operator and the patient by Means of Operator Protection (MOOP) and Means of Patient Protection (MOPP) classification in terms of the separation safety distance, insulation schemes, and dielectric strength requirements.
- The classification determined the mandated or allowed types of levels of isolation, insulation, creepage, air clearance, and leakage current that operators and patients may come into contact with.

#### ISOLATION VOLTAGE AND SAFETY DISTANCE IN THE 3rd EDITION OF THE IEC 60601-1 STANDARD

• The 3<sup>rd</sup> edition of the approved IEC 60601-1 safety standard specified that the safety distance for the minimum creepage, air clearance, and isolation voltage must be met in order to avoid risks and to ensure freedom from dangerous energy shocks due to any electrical shocks and excess energy hazards, transient voltage spikes, insulation breakdown of the power architecture, mechanical damage, ignition, fires, shorts developing between PCB tracks and across air gaps, arcing, and ground loops, which comply with the limited leakage current during normal and single-fault conditions.

The level requirement for the isolation voltage depends on the insulation type, the working voltage, and the degree of pollution, and the insulation barriers must have undergone a high-voltage test.

From our standpoint at MINMAX, we believe that the power supplies for medical and healthcare equipment should provide the highest degree of protection; therefore, our medical safety DC-DC converters and AC-DC power supplies have 2xMOPP/2xMOOP from the primary side to the secondary side (mains AC to low-voltage DC). This gives medical/healthcare equipment designers flexibility and assurance that there are primary and secondary reinforced insulation barriers in the medical power architecture to ensure long-term safety.

Insulation		МООР				
	Air Clearance	Creepage Distance	Test Voltage	Air Clearance	Creepage Distance	Test Voltage
Basic   1 x MOP	2.0 mm	3.2 mm	1500 VAC	2.5 mm	4.0 mm	1500 VAC
Double or Reinforced   2 x MOP	4.0 mm	6.4 mm	3000 VAC	5.0 mm	8.0 mm	4000 VAC

| Insulation test voltages based on 250 VAC working voltage | MOP = Means of protection | MOOP = Means of operation protection | | MOPP = Means of patient protection |

#### ISOLATION VOLTAGE AND SAFETY DISTANCE IN THE 3rd EDITION OF THE IEC 60601-1 STANDARD



#### LEAKAGE CURRENT IN THE 3rd EDITION OF THE IEC 60601-1 STANDARD

- Whether the product is considered MOOP or MOPP, the leakage-current requirements must be met.
- A further change between the 2<sup>nd</sup> and 3<sup>rd</sup> editions is related to the Earth leakage-current requirements.
- The leakage current needs to comply with a limit value to avoid risks and ensure freedom from dangerous energy shocks due to any electrical shock and excess energy hazards, transient voltage spikes, insulation breakdown of the power architecture, mechanical damage, ignition, fire, shorts developing between PCB tracks, air gaps, arcing, and ground loops that the operator or patient may come into direct contact during normal and single-fault conditions.
- Leakage-current tests are designed to simulate a human body coming into contact with different parts of the medical/healthcare equipment. The measured leakage-current values should comply with the acceptable limits.

E	Earth leakage current	Current flowing in the earth conductor.
Е	Enclosure leakage current	Current flowing to earth via the patient from the enclosure.
P	Patient leakage current	Current flowing to earth via the patient from an applied part.
P	Patient auxiliary current	Current flowing between two applied parts.

	TYPE	В	TYPE	BF	TYPE CF		
Leakage Current	NC	SFC	NC	SFC	NC	SFC	
Earth leakage current general	500 μA*	1000 μΑ	500 μ <b>A</b> *a	1000 μΑ*	500 μA*	1000 μΑ	
Enclosure leakage current	100 μΑ	500μA*	100μΑ	500μA*	100μΑ	500μA*	
Patient leakage current AC	100 μΑ	500μΑ	100μΑ	500μΑ	10μΑ	50μΑ	
Patient auxiliary current DC	10 μΑ	50μΑ	10μΑ	50μΑ	10μΑ	50μΑ	

| NC = Normal Conditions

| SFC = Single Fault Conditions|

 $\mid$  \*The maximum Earth and enclosure leakage current for patient care equipment in the US is 300  $\mu$ A.

[1] "Enclosure Leakage Current" changed to "Touch Leakage Current" in the 3rd edition of the IEC 60601-1 standard

#### **INSULATION TYPE**

• The five different types of insulation grades are listed below.

Functional Insulation	Insulation that is necessary only for the correct functioning of the equipment and does not provide any protection against electric shock.
Basic Insulation	Insulation applied to live parts to provide protection against electric shock.
Supplementary Insulation	Independent insulation applied in addition to basic insulation in order to provide protection against electric shock in the event of a failure of basic insulation.
Double Insulation	Insulation comprising both basic insulation and supplementary insulation.
Reinforced Insulation	Single insulation system applied to live parts which provide a degree of protection against electric shock equivalent to double insulation.


#### STRUCTURE OF THE MEDICAL INSULATION SYSTEM

- The figure below shows the insulation system structure and two insulation barriers that provide two Means of Protection (MOP) that must be present within medical/healthcare systems.
- Two insulation barriers are required to ensure that the applied part, which includes both patients and operators, is isolated and protected from dangerous energy shocks due to any electrical shocks and excess energy hazards, transient voltage spikes, insulation breakdown of the power architecture, mechanical damage, ignition, fire, shorts developing between PCB tracks, air gaps, arcing, and ground loops and complies with the limited leakage current during normal and single-fault conditions.
- · Within the electrical safety area, medical safety approval requires equipment to implement two Means of Protection (MOP) such that if a failure occurs within one area, a second mechanism safeguards the operator and/or patient against dangerous energy shocks from any electrical shocks and excess energy hazards, transient voltage spikes, and insulation breakdown of the power architecture.

MINMAX medical safety AC-DC power supplies and medical safety DC-DC converters have been approved to the new 3<sup>rd</sup> edition of the IEC 60601-1 standard with a reinforced insulation level. Further, all medical safety AC-DC power supplies provide 2xMOPP and are suitable as the primary reinforced insulation barrier. All medical safety DC-DC converters provide 2xMOPP or 2xMOOP and are suitable as the secondary reinforced insulation barrier of the insulation system structure to ensure the long-term safety of operators/patients.

#### STRUCTURE OF THE MEDICAL INSULATION SYSTEM

• Medical/Healthcare Insulation System



MP:

Live Part = Non Applied Part Electronic Circuit | LP Applied Part = Applied Part Electronic Circuit

#### **MEDICAL EMC STANDARDS**

- All MINMAX medical safety AC-DC power supplies and DC-DC converters undergo 4th edition medical EMC (Emission + Immunity) testing of LIFE-SUPPORTING ME EQUIPMENT to help us provide as much information as possible during the design-in process to ensure that medical/healthcare equipment comply with EN 50155: 2009+AI for EMI (Electromagnetic Interference) and EN 60601-1-2: 2015 for EMS (Electromagnetic Susceptibility).
- EMC Test Level of Medical Safety DC-DC Converters

	Phenomenon	EN 60601-1 Reference Clause	Reference Standard	MINMAX Test Level of Medical DC-DC Converters					
EMI					0.15 – 0.5MHz	Quasi-peak 79 dBuV	Average 66 dBuV		
	Conducted Emission	EN 55011	EN 55011	Group 1 Class A	0.5 – 5.0MHz	Quasi-peak 73 dBuV	Average 60 dBuV		
					5.0 – 30MHz	Quasi-peak 73 dBuV	Average 60 dBuV		
	De district Facilities	EN 55011	EN 55011	Group 1	30 – 230MHz	40 dBuV/m			
	Radiated Emission	EN 55011	EN 55011	Class A	230 – 1000MHz	47 dBuV/m			
	ESD Test	EN 60601-1-2	IEC 61000-4-2	Air Disch	narge: ±15KVDC	Contact Dischar	rge   ±8KVDC		
	Radiated Immunity (RS)	EN 60601-1-2	IEC 61000-4-3	80 to 270	0MHz : 10V/m   38	35 to 6000MHz   9	9-28 V/m		
	Electrical Fast Transient (EFT)	EN 60601-1-2	IEC 61000-4-4	L1, L2, L1	1+L2 : ±2KVDC				
EMS	Surge Immunity Test	EN 60601-1-2	IEC 61000-4-5	L1 to L2	: ±2KVDC				
	Conducted Immunity (CS)	EN 60601-1-2	IEC 61000-4-6	0.15 to 80MHz: 10Vrms   ISM Frequency: 6 Vrms					
	Power Frequency Magnetic Field (PFMF)	EN 60601-1-2	IEC 61000-4-8	X, Y, Z axi	is : 30 A/m				

#### MEDICAL EMC STANDARDS

• EMC Test Level of Medical Safety AC-DC Power Supplies

	Phenomenon	EN 60601-1 Reference Clause	Reference Standard	MINMAX Test Level of Medical DC-DC Converters						
	Conducted Emission				0.15 – 0.5MHz	Quasi-peak 66-56 dBuV	Average 56-46 dBuV			
EN41		EN 55011	EN 55011	Group 1 Class A	0.5 – 5.0MHz	Quasi-peak 56 dBuV	Average 46 dBuV			
EMI					5.0 – 30MHz	Quasi-peak 60 dBuV	Average 50 dBuV			
	Radiated Emission	EN 55044	EN EE044	Group 1	30 – 230MHz	30 dBuV/m				
	Radiated Emission	EN 55011	EN 55011	Class A	230 – 1000MHz	37 dBuV/m				
	ESD Test	EN 60601-1-2	IEC 61000-4-2	Air Disch	narge : ±15KVDC	Contact Dischar	ge   ±8KVDC			
	Radiated Immunity (RS)	EN 60601-1-2	IEC 61000-4-3	80 to 2700MHz : 10V/m   385 to 6000MHz   9-28 \						
	Electrical Fast Transient (EFT)	EN 60601-1-2	IEC 61000-4-4	L, N, L+N: ±2KVDC						
	Surge Immunity Test	EN 60601-1-2	IEC 61000-4-5	L to N:	±2KVDC					
	Conducted Immunity (CS)	EN 60601-1-2	IEC 61000-4-6	0.15 to 80	OMHz : 10Vrms   I	SM Frequency:	6 Vrms			
EMS	Power Frequency Magnetic Field (PFMF)	EN 60601-1-2	IEC 61000-4-8	X, Y, Z axis : 30 A/m						
	Voltage Dips	EN 60601-1-2	IEC 61000-4-11	Dips: 100% Reduction for 0.5 cycle at 50 Hz 100% Reduction for 1 cycle at 50 Hz 30% Reduction for 25/30 cycle at 50/60 Hz Interruptions: 100% Reduction for 250/300 cycle at 50/60 Hz						
	Short Interruptions	EIN OUOU I - I - Z	IEC 01000-4-11							

#### **COST VS. RISKS TO SAFETY**

- Although the 3<sup>rd</sup> edition of the IEC 60601-1 standard offers medical/healthcare equipment manufacturers more options related to the choice of power supply, the upcoming question of risk vs. cost must be considered, i.e., a cheaper power supply with a lower performance to save a few dollars versus a power supply with higher specifications and safety certification that might cost more but reduce the risk as much as possible. After all, if you get it wrong in medical device design, it can limit your market, compromise your brand, and severely delay regulatory approval or worse.
- As a result, the specifications of medical safety power supplies that are approved at the IEC/EN 60601-1 standard and comply with Means of Protection (MOP) is preferred for medical/healthcare equipment manufacturers.



### Fully Vacuum Encapsulated to Save Your System

• For the electromagnetic susceptibility and environmental physical stress interference which are protected from acomprehensive protection capabilities.




# Reinforced Insulation & 5KVAC Isolation for System Safety

• The 5KVAC I/O isolation with reinforced insulation and vacuum encapsulated creates a solid electrical barrier which to protect sensitive circuit load from noise, electromagnetic disturbances, power bus fluctuation, avoid the risk and ensure freedom dangerous energy shock from any electric shock and excess energy hazards, transient voltage spike, insulation breakdown of power architecture, mechanical damage, ignition, fire and short developing between PCB tracks, air gaps, arcing and ground loop that provide safety on long-term operation of medical/ healthcare equipment.

Please refer to the "Isolation Voltage and Safety Distance in the 3rd Edition of the IEC 60601-1 Standard & Leakage Current in the 3rd Edition of the IEC 60601-1 Standard & Insulation Type" on Page 04 & 05 & 06 for more information.

\* Example : MIW06-24S12M


## 8mm Creepage & Clearance with 2xMOPP Level



Isolation 5KVAC/60sec with Reinforced Insulation Low Leakage Current < 2uA Low I/O Isolation Capacitance 40pF max.

#### **Wider Operating Ambient Temperature Range**

• Wider operating temp. range by latest thermal management technology and fully vacuum encapsulated.





\* Example : MIW06-24S12M

#### ECO Technology

#### Green Design for Higher Full Range Efficiency

 High efficiency for whole output load, input line & ambient temp. range by latest green design technology helps to energy saving, thermal management, minimize the temp. rise and size miniaturization.








# **Green Design for Energy Saving, Minimize Temperature Rise**

 Ultra low no-load power consumption by latest green design technology helps to improve and minimize the temp. rise (avoid thermal problem), energy saving and prolong the battery life.

Very Low No Load Input Current 5mA @ 24Vin Very Low No Load Power Consumption 0.12Watt @ 24Vin



# Green Design for No Min. Load / Dummy Load Requirement

 With high stability feedback loop design, the MINMAX medical safety power solutions may not oscillate in no-load or light-load condition.



# MINMAX TECHNOLO

#### **Power Your System Precisely**


- Setting Accuracy ±1% Vom | Line Regulation ±0.5% | Load Regulation ±0.5% | High Transient Response | Low Temperature Coefficient |
- The output voltage will still keep with excellent accuracy, even though the input voltage, output current and ambient temp. of the system are unstable.
- The output voltage of mainstream products may undershoot and overshoot obviously during the load changes.
- MINMAX medical safety power solutions are still keep with rated output voltage preciously.




#### **Faster Start-up Time without Overshoot**


 The start-up time of MIW06M decreases from 250mS to 20mS which helps to avoid any system timing failure caused by long start-up time.

Faster start-up time without overshoot ensures the safety of your system.



Example: MIW06-24S12M






Example: Mainstream/Before


#### **Superior Load Driving Capability**


 MINMAX medical safety power solutions have superior load driving capability which can drive your system during very low voltage and even zero voltage output without start-up failure.




#### **Excellent EMC Performance**


- Excellent EMI performance by upgraded noise filtering technology helps to improve overall system EMI performance on conduction and radiation emission.
- No external component needed for conducted emission meets EN 55011 Class A.








- Excellent EMS performance by upgraded noise immunity technology helps to improve overall system EMS performance on ESD, Surge, EFT, RS, CS and PFMF.
- Only one E-cap. needed for ±2KV surge immunity by IEC 61000-4-5 with criteria A.




 Only one E-cap. needed for ±2KV EFT immunity by IEC 61000-4-4 with criteria A.



<sup>\*</sup> Please refer to the "Medical EMC Standards" on Page 09 & 10 for more information.


#### Lower Ripple & Noise

 Small Ripple & Noise for whole output load, input line & ambient temp. range by upgraded noise filtering technology helps to reduce the peripheral components needed and noise interference.



• MINMAX medical safety power solutions are the instantaneous protection for UVP, OCP, OVP & SCP which help to ensure long-term safety of system operation.

**Protection Functions on Abnormal Operation** 



\*Example : MIW06-24S12M

#### Certifications

- **UL/cUL/IEC/EN 62368-1(60950-1) Approved**
- **UL** 508 Approved
- Medical Safety IEC/EN 60601-1 & ANSI/AAMI ES 60601-1 Approved
- **CE Marking**
- Protection Class



## MINMAX MEDICAL SAFETY POWER SOLUTIONS

#### **SUCCESSFUL APPLICATIONS**





• Mobile Cart Computer



• CT Scanning



Medical Panel PC



- ECG, EEG
- Medical Assist System
- CPAP Machine



• Healthcare Information System



#### **MEDICAL SAFETY PRODUCTS OVERVIEW**

#### Medical Safety DC-DC Converters

Medical	Jaic	ty DC-D	CCOIIV	ei tei 5										
Series	Output Power	Input Voltage Range (VDC)	Output Voltage (VDC)	Isolation (VDC)	Efficiency	Operating Ambient Temp. Range	Output Regulation	No Min. Load					ANSI/AAMI ES 60601-1	
1W • SIP Pag	ckage													
MAU400	1W	4.5-5.5, 10.8-13.2	5, 12, 15, ±12, ±15	3000VAC Reinforced	75%	-25~+85°C						1xM0PP 2xM00P	•	•
MAU01M	1W	4.5-5.5, 10.8-13.2 21.6-26.4,	5, 12, 15	4000VAC Reinforced	81%	-40~+95°C					•	2xM0PP	•	•
1-2W • SMD	Packa	ge												
MSCU01M	1W	4.5-5.5, 10.8-13.2, 21.6-26.4	5, 12, 15, ±12, ±15	4000VAC Reinforced	84%	-40~+95°C					•	2xM0PP	•	•
MSHU100	2W	4.5-5.5, 10.8-13.2 21.6-26.4,	5, 12, 15, ±12, ±15	4000VAC Reinforced	75%	-25~+80°C						1xM0PP 2xM00P	•	•
2-10W • DIP	Packa	ge												
MDHU100	2W	4.5-5.5, 10.8-13.2 21.6-26.4,	5, 12, 15, ±12, ±15	4000VAC Reinforced	75%	-25~+80°C						1xM0PP 2xM00P	•	
MIHW2000	3W	9-40, 18-80, 36-160	5, 12, ±12, ±15	4000VAC Reinforced	83%	-40~+85°C	•		•		•	1xM0PP 2xM00P	•	•
MIW03M	3.5W	4.5-9, 9-18, 18-36, 36-75	5, 5.8, 12, 15 ±12, ±15	5000VAC Reinforced	87%	-40~+96°C	٠	•	•		•	2xM0PP	•	
MIW06M	6W	9-18, 18-36, 36-75	5, 12, 15, ±12, ±15	5000VAC Reinforced	89%	-40~+95°C	•	•	•	•	•	2xM0PP	•	•
MIW10M	10W	9-18, 18-36, 36-75	3.3, 5, 5.1, 12, 15, 24, ±12, ±15	5000VAC Reinforced	88%	-40~+90°C	•	۰	•	•	•	2xM0PP	•	
10-20W • 2"	×1" Pa	ckage												
MKW15M	15W	9-18, 18-36, 36-75	5, 5.1, 12, 15, 24, ±12, ±15	4200VAC Reinforced	90%	-40~+85°C	•	۰	•	•		2xM0PP	•	•
MKW20M	20W	9-18, 18-36, 36-75	5, 5.1, 12, 15, 24, ±12, ±15	4200VAC Reinforced	90%	-40~+80°C	٠	۰	•	•	•	2xM0PP	•	

#### Medical Safety AC-DC Converters

Series	Output Power	Input Voltage Range (VAC)	Output Voltage (VDC)	Isolation (VDC)	Efficiency	Operating Ambient Temp. Range	Output Regulation	No Min. Load	Package	OCP	OVP		ANSI/AAMI ES 60601-1	1EC/EN 60601-1 3 rd
AJM-24	24W	85-264	5,9,12,15, 24,±12 ±15	4000VAC Reinforced	85%	-40~+80°C	•	•	PCB Chassis DIN-Rail	•	•	2xM0P	P •	•
APM-40	40W	85-264	5,12,15, ±12±15	4000VAC Reinforced	85%	-40~+80°C	۰	•	PCB Chassis DIN-Rail	•	•	2xM0P	P •	•
AYM-60	60W	85-264	5.1,12, 15,24,48	4000VAC Reinforced	85%	-40~+80°C	•	•	PCB Chassis DIN-Rail	•	•	2xM0P		•

<sup>&</sup>lt;sup>11</sup>Please refer to derating curve information form datasheet

#### **MEDICAL SAFETY • DC-DC CONVERTERS**

MAU400 Series SIP Package











Model Selection Guide									
Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency					
MAU401		5	200	66%					
MAU402		12	80	66%					
MAU403	4.5 - 5.5	15	65±	66%					
MAU404	4.0 0.0	±5	100	66%					
MAU405		±12	±40	72%					
MAU406		±15	±35	73%					
MAU411		5	200	66%					
MAU412		12	80	66%					
MAU413	10.8 - 13.2	15	65	66%					
MAU414		±5	±100	66%					
MAU415		±12	±40	74%					
MAU416		±15	±35	75%					

MAU01M Series SIP Package





79%

79%



	Input Voltage (VDC)	Output Voltage (VDC)		Efficiency
/AU01-05S05M		5	200	79%
/AU01-05S12M	4.5 - 5.5	12	84	80%
4AU01-05S15M		15	68	81%
4AU01-12S05M		5	200	79%
4AU01-12S12M	10.8 - 13.2	12	84	81%
MAU01-12S15M		15	68	79%
4AU01-24S05M		5	200	76%

MAU01-24S12M

MAU01-24S15M

**Model Selection Guide** 

\*There are different features & spec. by each series. For detailed series datasheet, please refer to www.minmaxpower.com

#### **MEDICAL SAFETY • DC-DC CONVERTERS**

#### MSCU01M Series SMD Package







Model Selection Guide										
Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency						
MSCU01-05S05M		5	200	76%						
MSCU01-05S12M		12	84	80%						
MSCU01-05S15M	4.5 - 5.5	15	68	83%						
MSCU01-05D12M		±12	±42	80%						
MSCU01-05D15M		±15	±33	84%						
MSCU01-12S05M		5	200	76%						
MSCU01-12S12M		12	84	79%						
MSCU01-12S15M	10.8 - 13.2	15	68	80%						
MSCU01-12D12M		±12	±42	79%						
MSCU01-12D15M		±15	±33	80%						
MSCU01-24S05M		5	200	76%						
MSCU01-24S12M		12	84	80%						
MSCU01-24S15M	21.6 - 26.4	15	68	80%						
MSCU01-24D12M		±12	±42	80%						
MSCU01-24D15M		±15	±33	80%						

#### MSHU100 Series SMD Package











Model Selection Guide						
Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency		
MSHU102		5	400	66%		
MSHU104	4.5 - 5.5	12	165	66%		
MSHU105	4.0 0.0	15	133	66%		
MSHU108		±12	±83	72%		
MSHU109		±15	±66	73%		
MSHU112		5	400	66%		
MSHU114		12	165	66%		
MSHU115	10.8 - 13.2	15	133	66%		
MSHU118		±12	±83	74%		
MSHU119		±15	±66	75%		
MSHU122		5	400	66%		
MSHU124		12	165	66%		
MSHU125	21.6 - 26.4	15	133	66%		
MSHU128		±12	±83	74%		
MSHU129		±15	±66	75%		

\*There are different features & spec. by each series. For detailed series datasheet, please refer to www.minmaxpower.com

#### **MEDICAL SAFETY • DC-DC CONVERTERS**

#### MDHU100 Series DIP Package













Model Number	Input Voltage (VDC)	Output Voltage (VDC)		
MDHU102		5	400	66%
MDHU104		12	165	66%
MDHU105	4.5 - 5.5	15	133	66%
MDHU108		±12	±83	72%
MDHU109		±15	±66	73%
MDHU112		5	400	66%
MDHU114		12	165	66%
MDHU115	10.8 - 13.2	15	133	66%
MDHU118		±12	±83	74%
MDHU119		±15	±66	75%
MDHU122		5	400	66%
MDHU124		12	165	66%
MDHU125	21.6 - 26.4	15	133	66%
MDHU128		±12	±83	74%
MDHU129		±15	±66	75%

#### MIHW2000 Series DIP Package







#### **Model Selection Guide**



Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency
MIHW2022		5	600	78%
MIHW2023	9 - 40	12	250	83%
MIHW2026	7 - 40	±12	±125	83%
MIHW2027		±15	±100	83%
MIHW2032		5	600	78%
MIHW2033	18 - 80	12	250	83%
MIHW2036	10 00	±12	±125	83%
MIHW2037		±15	±100	83%
MIHW2042		5	600	78%
MIHW2043		12	250	83%
MIHW2046	36 - 160	±12	±125	83%
MIHW2047		±15	±100	83%
	*Thora	are different	fasturas & spac	by each carios

\*There are different features & spec. by each series. For detailed series datasheet, please refer to www.minmaxpower.com

#### **MEDICAL SAFETY • DC-DC CONVERTERS**

#### MIW03M Series DIP Package









Model Selection Guide					
	Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency
	MIW03-05S05M MIW03-05S058M MIW03-05S12M MIW03-05S15M MIW03-05D12M MIW03-05D15M	4.5 - 9	5 5.8 12 15 ±12 ±15	700 600 290 235 ±145 ±115	83% 83% 84% 84% 84%
	MIW03-12S05M MIW03-12S12M MIW03-12S15M MIW03-12D12M MIW03-12D15M	9 - 18	5 12 15 ±12 ±15	700 290 235 ±145 ±115	83% 87% 87% 87% 87%
	MIW03-24S05M MIW03-24S12M MIW03-24S15M MIW03-24D12M MIW03-24D15M	18 - 36	5 12 15 ±12 ±15	700 290 235 ±145 ±115	83% 86% 87% 87% 86%
	MIW03-48505M MIW03-48512M MIW03-48515M MIW03-48D12M MIW03-48D15M	36 - 75	5 12 15 ±12 ±15	700 290 235 ±145 ±115	83% 86% 85% 84% 84%

#### MIW06M Series DIP Package







Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency
MIW06-12S05M		5	1200	84%
MIW06-12S12M		12	500	87%
MIW06-12S15M	9 - 18	15	400	86%
MIW06-12D12M		±12	±250	87%
MIW06-12D15M		±15	±200	87%
MIW06-24S05M		5	1200	84%
MIW06-24S12M		12	500	87%
MIW06-24S15M	18 - 36	15	400	87%
MIW06-24D12M		±12	±250	86%
MIW06-24D15M		±15	±200	87%
MIW06-48S05M		5	1200	84%
MIW06-48S12M		12	500	87%
MIW06-48S15M	36 - 75	15	400	89%
MIW06-48D12M		±12	±250	87%
MIW06-48D15M		±15	±200	88%

Model Selection Guide

\*There are different features & spec. by each series. For detailed series datasheet, please refer to www.minmaxpower.com

#### **MEDICAL SAFETY • DC-DC CONVERTERS**

#### MIW10M Series DIP Package









Model Number	Input Voltage (VDC)	Output Voltage (VDC)		
MIW10-12S033M MIW10-12S05M MIW10-12S051M MIW10-12S12M MIW10-12S15M MIW10-12S24M MIW10-12D12M MIW10-12D15M	9 - 18	3.3 5 5.1 12 15 24 ±12 ±15	2700 2000 2000 833 666 416 ±416 ±333	81% 84% 84% 87% 88% 88% 88%
MIW10-24S033M MIW10-24S05M MIW10-24S051M MIW10-24S12M MIW10-24S15M MIW10-24S24M MIW10-24D12M MIW10-24D15M	18 - 36	3.3 5 5.1 12 15 24 ±12 ±15	2700 2000 2000 833 666 416 ±416 ±333	81% 85% 85% 88% 88% 88% 88%
MIW10-48S033M MIW10-48S05M		3.3 5	2700 2000	81% 85%

**Model Selection Guide** 

#### MKW15M Series 2"x1" Package

MIW10-48S051M MIW10-48S12M MIW10-48S15M

MIW10-48S24M

MIW10-48D12M MIW10-48D15M





416

±416



88% 87% 87%



Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency
MKW15-12S05M MKW15-12S051M MKW15-12S12M MKW15-12S12M MKW15-12S24M MKW15-12D12M MKW15-12D15M	9 - 18	5 5.1 12 15 24 ±12 ±15	3000 3000 1250 1000 625 ±625 ±500	86% 86% 89% 88% 88% 88%
MKW15-24S05M MKW15-24S051M MKW15-24S12M MKW15-24S15M MKW15-24S15M MKW15-24S24M MKW15-24D12M MKW15-24D15M	18 - 36	5 5.1 12 15 24 ±12 ±15	3000 3000 1250 1000 625 ±625 ±500	88% 88% 89% 89% 90% 90%
MKW15-48505M MKW15-485051M MKW15-48512M MKW15-48515M MKW15-48524M MKW15-48D12M MKW15-48D15M	36 - 75	5 5.1 12 15 24 ±12 ±15	3000 3000 1250 1000 625 ±625 ±500	88% 88% 88% 90% 89% 89% 88%

**Model Selection Guide** 

\*There are different features & spec. by each series. For detailed series datasheet, please refer to www.minmaxpower.com

#### **MEDICAL SAFETY • DC-DC CONVERTERS**

MKW20M Series 2"x1" Package







	Model Selection Guide					
	Model Number	Input Voltage (VDC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency	
	MKW20-12S05M MKW20-12S051M MKW20-12S12M MKW20-12S15M MKW20-12S24M MKW20-12D12M MKW20-12D15M	9 - 18	5 5.1 12 15 24 ±12 ±15	4000 4000 1670 1333 840 ±840 ±670	86% 86% 89% 88% 89% 89%	
	MKW20-24S05M MKW20-24S051M MKW20-24S12M MKW20-24S15M MKW20-24S24M MKW20-24D12M MKW20-24D15M	18 - 36	5 5.1 12 15 24 ±12 ±15	4000 4000 1670 1333 840 ±840 ±670	88% 88% 89% 90% 90%	
	MKW20-48S05M MKW20-48S051M MKW20-48S12M MKW20-48S15M MKW20-48S24M MKW20-48D12M MKW20-48D15M	36 - 75	5 5.1 12 15 24 ±12 ±15	4000 4000 1670 1333 840 ±840 ±670	88% 88% 89% 90% 89% 89%	

\*There are different features & spec. by each series. For detailed series datasheet, please refer to www.minmaxpower.com

#### **MEDICAL SAFETY • AC-DC POWER SUPPLIES**

AJM-24 Series Encapsulated Package

ANSI/AAM CANSI/AAM C











Model Selection			Selection Guid	de
	Model Number	Input Voltage (VAC)	Output Voltage (VDC)	Output Curren (mA)ma
	AJM-24S05		5	3000
	AJM-24S09		9	2666
	AJM-24S12		12	2000
	AJM-24S15	85 - 264	15	1600
	AJM-24S24		24	1000
	AJM-24D12		±12	±1000

APM-40 Series Encapsulated Package

AJM-24D15



±15





±800



77% 82% 83% 82%

85% 84%

84%



Model Selection Guide					
	Model Number	Input Voltage (VAC)	Output Voltage (VDC)	Output Current (mA)max	Efficiency
	APM-40S05		5	8000	81%
	APM-40S12		12	3330	84%
	APM-40S15		15	2660	85%
	APM-40S24	85 - 264	24	1660	84%
	APM-40D12		±12	±1660	84%
	APM-40D15		±15	±1330	85%

AYM-60 Series Encapsulated Package









MINMAX TECHNOLOGY



Model	Sel	lection	Guide

Model Number		Output Voltage (VDC)		
AYM-60S051		5.1	10000	84%
AYM-60S12		12	5000	87%
AYM-60S15	85 - 264	15	4000	87%
AYM-60S24		24	2500	87%
AYM-60S48		48	1250	88%



• DC-DC CONVERTERS


• AC-DC POWER SUPPLIES



40-50W



# MINMAX POWER SOLUTIONS DC-DC CONVERTERS · AC-DC POWER SUPPLIES | 1-150W



DC-DC CONVERTERS





6-60W

• DC-DC CONVERTERS


10-30W





• DC-DC CONVERTERS

2-10W



15-20W





